Progress report

2018/11/09
Michitaka Inomoto
Tokyo University of Science

What I have done

I finished improving the setup.

Then, I'm measuring PMT's efficiency in B-field with the coil.

The coil makes B-field: about 0.9mT in 2A.



03180529

LED fiber and fixture by 3D printer

interior of dark box