

Automated Calibration System for the Daya Bay Experiment




Jianglai Liu


Shanghai Jiao Tong University

On behalf of the Daya Bay Collaboration

Daya Bay experiment overview






To first order



Anti-neutrino detector (AD)

Inverse Beta Decay



Cylindrical 3-zone Structure
Separated By Acrylic Vessels:

I. Target: 0.1% Gd-loaded liquid scintillator, 20 ton

II. Gamma-catcher: liquid scintillator,

III. Buffer shielding: mineral oil

Acrylic vessel thickness: 1.5 cm (outer) and 1 cm (inner)

192 8" PMT's on circumference and reflective reflectors on top and bottom.

6 'functionally identical' detectors:
Reduce systematic uncertainties

$$\frac{N_f}{N_n} = \left(\frac{N_{p,f}}{N_{p,n}} \right) \left(\frac{L_n}{L_f} \right)^2 \left(\frac{\epsilon_f}{\epsilon_n} \right) \left[\frac{P_{\text{sur}}(E, L_f)}{P_{\text{sur}}(E, L_n)} \right]$$

Design principle of the automated calibration system


$$\frac{N_f}{N_n} = \left(\frac{N_{p,f}}{N_{p,n}} \right) \left(\frac{L_n}{L_f} \right)^2 \left(\frac{\varepsilon_f}{\varepsilon_n} \right) \frac{P_{survival}(E, L_f)}{P_{survival}(E, L_n)} \rightarrow \sin^2 2\theta_{13}$$

1? Key target of the calibration program


- | | |
|---|--|
| Routine (weekly) deployment of sources | → Automated calibration unit |
| No introduction of Rn background | → ACU stays with AD |
| Simple, robust, and minimize material in liquid | → z scan only |
| Sample target and gamma catcher region | → Multiple unit on each AD |
| LED light = fix time | → Timing and gain calibration |
| Radioactive source = fix energy | → e ⁺ and neutron sources for energy and efficiency calibration |

Overview

R=1.7725 m R=0 R=1.35m




Three axes: center, edge of target, middle of gamma catcher



3 sources for each z axis on a turntable (position accuracy < 5 mm):

- 10 Hz ^{68}Ge
- 0.5 Hz ^{241}Am - ^{13}C neutron source + 100 Hz ^{60}Co gamma source
- LED diffuser ball (500 Hz)


Mechanical construction




Compatibility with LAB-based scintillator: teflon, acrylic (enclosed), stainless steel

Gamma sources

- Ge-68 (Rate: 10 Bq, $T_{1/2}$: 270 days)



- Positron threshold
- Relative PMT detecting efficiency
- Co-60 (Rate: 150 Bq, $T_{1/2}$: 1925 days)




- Energy calibration
- Monitor light yield/attenuation


5 x 5 mm



Neutron source: AmC
(Next slide)



Neutron source






^{241}Am - ^{13}C (Rate: 0.5 Hz)


- $^{13}\text{C} (\alpha, n) ^{16}\text{O}$
- Au foil attenuate α to < 4.5 MeV, hence suppressing 6.13 MeV gamma from excited state of ^{16}O .
- Neutron energy scale



Light source



Interface with the AD




- Source port in ACU aligned with acrylic vessel penetrations
- Source deploy through the penetration by gravity
- Absolute position accuracy – 5 mm

Control and software

- Control software written with LabVIEW.
- Handshake between control software and DAQ → complete automated calibration and synchronized data taking



40x speed




Operate ACU

```
1 <DeploymentScript>
2   <ActivateACU AcuNumber="1"/>
3   <Move Depth="2" MovementType="M0" SourceNumber="1"/>
4   <PulseLED Frequency="500" Handshake="True" LTBMode="Fals
5   <Move Depth="1.9" MovementType="M0" SourceNumber="1"/>
6   <PulseLED Frequency="500" Handshake="True" LTBMode="Fals
7   <Move Depth="1.8" MovementType="M0" SourceNumber="1"/>
8   <PulseLED Frequency="500" Handshake="True" LTBMode="Fals
9   <Move Depth="1.7" MovementType="M0" SourceNumber="1"/>
10  <PulseLED Frequency="500" Handshake="True" LTBMode="Fals
11  <Move Depth="1.6" MovementType="M0" SourceNumber="1"/>
12  <PulseLED Frequency="500" Handshake="True" LTBMode="Fals
13  <Move Depth="1.5" MovementType="M0" SourceNumber="1"/>
14  <PulseLED Frequency="500" Handshake="True" LTBMode="Fals
15  <Move Depth="1.4" MovementType="M0" SourceNumber="1"/>
16  <PulseLED Frequency="500" Handshake="True" LTBMode="Fals
```

Input XML script





Calibration system in experimental halls





Example analysis: energy calibration

Weekly deployments of ^{60}Co at detector center: Monitor photoelectrons collected per MeV



3 sources along 3 axes



Summary of features

- Completely automated
- Three movable units per detector, stayed on the detector lid, accessing three vertical axis in the detector
- LED, gamma, and neutron sources in each unit
- Stringent material selection and cleanliness requirement met
- Under smooth operation since the start of data taking (Dec. 2011).

All design goals fulfilled